



California Medical Device Recall Information



Recall Name

CareFusion Recalls Alaris PC Unit (Model 8015) Due To Possible Malfunction

Recall Date	Product Description	Recalling Firm	Recall Reason
6/17/13	Alaris PC unit with Version 9.12 Keyboard Processor	CareFusion Corp. San Diego, CA	<i>Could potentially experience a loss of communication between the PC Unit main Processor and Keyboard Processor.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Alaris PC unit (Model 8015) with software version 9.12 • Affected Serial Number List	CA , nationwide	Distributed: September 21, 2012 through May 17, 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm364264.htm>